

Radioactive Material Inventory and Leak Test Form

36th Civil Engineering Squadron (EM)

Permit #: **Generally Licensed Material**

Inventory Frequency: Quarterly Semi-Annual Annual Other _____

Leak Test Frequency: Quarterly Semi-Annual Annual Other _____



Source Information

Isotope	Activity	Item Description	Unit Serial #	Source Serial #	Storage Location
---------	----------	------------------	---------------	-----------------	------------------

Leak Test Information¹

Swipe Sample #	Date Collected	Result
----------------	----------------	--------

Ni-63	10 mCi	APD 2000	5691		CEX Warehouse	WW120079	12-Jun-13	
Ni-63	10 mCi	APD 2000	5690		CEX Warehouse	WW120076	12-Jun-13	
Ni-63	10 mCi	APD 2000	5692		CEX Warehouse	WW120077	12-Jun-13	
Ni-63	10 mCi	APD 2000	3123		CEX Warehouse	WW120078	12-Jun-13	

¹Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

All items accounted for
 All swipe results less than or equal to 0.005 uCi

Date: 12 Jun 13
Date Results Received:

Survey Accomplished By: _____ **Reviewed By:** _____

Name: **KRISTIAN M. ATKINSON, SSgt, USAF, 4B071** **KHAI H. VUONG, Maj, USAF, BSC**

Signature: _____