

**Radioactive Material Inventory and Leak Test Form**

**36th Civil Engineering Squadron (EM)**

**Permit #:** **Generally Licensed Material**

---

**Inventory Frequency:**        Quarterly     Semi-Annual     Annual     Other \_\_\_\_\_

---

**Leak Test Frequency:**        Quarterly     Semi-Annual     Annual     Other \_\_\_\_\_



**Source Information**

**Leak Test Information<sup>1</sup>**

Isotope	Activity	Item Description	Unit Serial #	Source Serial #	Storage Location	Swipe Sample #	Date Collected	Result
Ni-63	10 mCi	APD 2000	5691		CEX Warehouse	WW120062	11-Dec-12	
Ni-63	10 mCi	APD 2000	5690		CEX Warehouse	WW120059	11-Dec-12	
Ni-63	10 mCi	APD 2000	5692		CEX Warehouse	WW120061	11-Dec-12	
Ni-63	10 mCi	APD 2000	3123		CEX Warehouse	WW120060	11-Dec-12	

<sup>1</sup>Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

All items accounted for

All swipe results less than or equal to 0.005 uCi

Date: 11 Dec 12

Date Results Received: \_\_\_\_\_

<b>Survey Accomplished By:</b>	<b>Reviewed By:</b>
Name: <b>KRISTIAN M. ATKINSON, SSgt, USAF, 4B071</b>	<b>KHAI H. VUONG, Maj, USAF, BSC</b>
Signature: _____	_____