

Radioactive Material Inventory and Leak Test Form

Postal Service Center

Permit #: **Generally Licensed Material**

Inventory Frequency: Quarterly Semi-Annual Annual Other _____

Leak Test Frequency: Quarterly Semi-Annual Annual Other _____

Source Information						Leak Test Information ¹		
Isotope	Activity	Item Description	Unit Serial #	Source Serial #	Storage Location	Swipe Sample #	Date Collected	Result
Ni-63	10 mCi	IonScan 400	13143		Bldg 21001, Post Office	WW120075	12-Jun-13	

¹Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

- All items accounted for
 All swipe results less than or equal to 0.005 uCi
- Date: 12 Jun 13
Date Results Received:

Survey Accomplished By: **KRISTIAN M. ATKINSON, SSgt, USAF, 4B071** Reviewed By: **KHAI H. VUONG, Maj, USAF, BSC**

Name: _____

Signature: _____