

Radioactive Material Inventory and Leak Test Form

Postal Service Center

Permit #: **Generally Licensed Material**

Inventory Frequency: Quarterly Semi-Annual Annual Other _____

Leak Test Frequency: Quarterly Semi-Annual Annual Other _____

Source Information

Leak Test Information¹

Isotope	Activity	Item Description	Unit Serial #	Source Serial #	Storage Location	Swipe Sample #	Date Collected	Result
Ni-63	10 mCi	IonScan 400	13143		Bldg 21001, Post Office	WW-131209	17-Dec-13	

¹Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

All items accounted for
 All swipe results less than or equal to 0.005 uCi

Date: 23 Dec 13
Date Results Received:

Survey Accomplished By:
Name: **JEREMY A. BOLDT, SSgt, USAF, 4B071**

Reviewed By:
GREGORY J. LATHBURY, Capt, USAF, BSC

Signature:

Signature: