Radioactive Material Inventory and Leak Test Form

Postal Service Center

Permit #: Generally Licensed Material

Inventory Frequency: ☐ Quarterly ☑ Semi-Annual ☐ Annual ☐ Other _____

Leak Test Frequency: ☐ Quarterly ☑ Semi-Annual ☐ Annual ☐ Other _____

Source Information						Leak Test Information[1]		
Isotope	Activity	Item Description	Unit Serial #	Source Serial #	Storage Location	Swipe Sample #	Date Collected	Result
Ni-63	10 mCi	IonScan 400	13143		Bldg 21001, Post Office	WW-131209	17-Dec-13	

[1]Leak tests are not required for beta/gamma source with activity less than 100 uCi; nor for alpha sources less than 10 uCi.

☑ All items accounted for

☐ All swipe results less than or equal to 0.005 uCi

Date: 13 Jun 14

Date Results Received:

Survey Accomplished By:	Reviewed By:
Name: JEREMY A. BOLDT, SSgt, USAF, 4B071	GREGORY J. LATHBURY, Capt, USAF, BSC
Signature:	